

## § 486.102

in conformity with Federal, State, and local laws relating to safety standards.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

### § 486.102 Condition for coverage: Supervision by a qualified physician.

Portable X-ray services are provided under the supervision of a qualified physician.

(a) *Standard—physician supervision.* The performance of the roentgenologic procedures is subject to the supervision of a physician who meets the requirements of paragraph (b) of this section and one of the following requirements is met:

(1) The supervising physician owns the equipment and it is operated only by his employees, or

(2) The supervising physician certifies annually that he periodically checks the procedural manuals and observes the operators' performance, that he has verified that equipment and personnel meet applicable Federal, State, and local licensure and registration requirements and that safe operating procedures are used.

(b) *Standard—qualifications of the physician supervisor.* Portable X-ray services are provided under the supervision of a licensed doctor of medicine or licensed doctor of osteopathy who is qualified by advanced training and experience in the use of X-rays for diagnostic purposes, i.e., he (1) is certified in radiology by the American Board of Radiology or by the American Osteopathic Board of Radiology or possesses qualifications which are equivalent to those required for such certification, or (2) is certified or meets the requirements for certification in a medical specialty in which he has become qualified by experience and training in the use of X-rays for diagnostic purposes, or (3) specializes in radiology and is recognized by the medical community as a specialist in radiology.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

## 42 CFR Ch. IV (10–1–99 Edition)

### § 486.104 Condition for coverage: Qualifications, orientation and health of technical personnel.

Portable X-ray services are provided by qualified technologists.

(a) *Standard—qualifications of technologists.* All operators of the portable X-ray equipment meet the requirements of paragraph (a) (1), (2), or (3) of this section:

(1) Successful completion of a program of formal training in X-ray technology of not less than 24 months' duration in a school approved by the Council on Education of the American Medical Association or by the American Osteopathic Association, or have earned a bachelor's or associate degree in radiologic technology from an accredited college or university.

(2) For those whose training was completed prior to July 1, 1966, but on or after July 1, 1960: Successful completion of 24 full months of training and/or experience under the direct supervision of a physician who is certified in radiology by the American College of Radiology or who possesses qualifications which are equivalent to those required for such certification, and at least 12 full months of pertinent portable X-ray equipment operation experience in the 5 years prior to January 1, 1968.

(3) For those whose training was completed prior to July 1, 1960: Successful completion of 24 full months of training and/or experience of which at least 12 full months were under the direct supervision of a physician who is certified in radiology by the American College of Radiology or who possesses qualifications which are equivalent to those required for such certification, and at least 12 full months of pertinent portable X-ray equipment operation experience in the 5 years prior to January 1, 1968.

(b) *Standard—personnel orientation.* The supplier of portable X-ray services has an orientation program for personnel, based on a procedural manual which is: Available to all members of the staff, incorporates relevant portions of professionally recognized documents, and includes instruction in all of the following: